



Editorial: Biomacromolecules Systems Applied to Medical Implants for the Release of Therapeutic Agents

Joao Henrique Lopes¹, Nicolas Tabary² and Jacobo Hernandez-Montelongo^{3*}

¹Department of Chemistry, Division of Fundamental Sciences (IEF), Aeronautics Institute of Technology - ITA, Sao Jose dos Campos, Brazil, ²Université Lille 1, Unité Matériaux et Transformations (UMET), UMR CNRS 8207, Villeneuve d'Ascq, France, ³Department of Physical and Mathematical Sciences, Catholic University of Temuco, Temuco, Chile

Keywords: biomacromolecules, medical implants, controlled release, therapeutic agents, drug release

Editorial on the Research Topic

Biomacromolecules Systems Applied to Medical Implants for the Release of Therapeutic Agents

Medical implants are devices fabricated to replace, aid, or enhance some missing biological structure (Singh et al., 2020). These devices include catheters, stents, hip joints, intraocular lenses, and others. The implant surfaces are designed with biocompatible materials since they will be in direct contact with physiological fluids and tissues, such as titanium, silicone, bioactive glass, polymers, among others. In addition, the choice of the appropriate biomaterial is guided by the presence of desirable properties according to the claimed application (Ratner and Zhang, 2020). Surgery is the most common way to incorporate medical implants into the body, which involves inherent risks: bacterial infections, local swelling and induration, inadequate healing, triggering immune reactions, and other different possible complications (Major et al., 2015). The eventuality and intensity of these postoperative risks are higher in cases of immunosuppression caused by certain diseases or conditions, such as AIDS, cancer, diabetes, malnutrition, and certain genetic disorders (Madsen et al., 2021). In that sense, biomacromolecules are versatile to be configured as release systems of therapeutic agents to health or significantly reduce those postoperative risks.

This Research Topic collection covers the use of biomacromolecules systems applied to biomedical implants for the release of therapeutic agents. This collection of papers is formed of seven original research articles and one review article.

Different interesting antibiotic drug delivery systems based on biomacromolecules were presented in this Research Topic. New proposes were physiochemically detailed, their biocompatibility was confirmed and their antibacterial activity was tested against both Gram-negative and Gram-positive bacteria. Mansouri et al. obtained an optimal niosomal formulation that exhibited controlled drug release for the antibiotic streptomycin sulfate, with the highest encapsulation efficiency but a minimum size and low polydispersity index. Sun et al. explored a novel temperature-sensitive hydrogel-based on a dual sustained-release system prepared by mesoporous silica (SBA-15) loaded with the antibiotic vancomycin hydrochloride (Van). These Van/SBA-15 particles were encapsulated in a chitosan-glycerophosphate-sodium alginate hydrogel (CS-GP-SA) showing the controlled release of VAN for the treatment of infectious jaw defects. In the same field, Wiggers et al. obtained chitosan films crosslinked with quercetin for the controlled delivery of the antibiotic trimethoprim. Their results showed that quercetin increased the crosslinking degree in a dose-dependent manner, thus influencing the release kinetic of the loaded antibiotic while maintaining its bactericidal effects.

OPEN ACCESS

Edited and reviewed by:

Hasan Uludag,
University of Alberta, Canada

*Correspondence:

Jacobo Hernandez-Montelongo
jacobo.hernandez@uct.cl

Specialty section:

This article was submitted to
Biomaterials,
a section of the journal
Frontiers in Bioengineering and
Biotechnology

Received: 01 April 2022

Accepted: 12 April 2022

Published: 27 April 2022

Citation:

Lopes JH, Tabary N and
Hernandez-Montelongo J (2022)
Editorial: Biomacromolecules Systems
Applied to Medical Implants for the
Release of Therapeutic Agents.
Front. Bioeng. Biotechnol. 10:910203.
doi: 10.3389/fbioe.2022.910203

Boas and Rechtes reported the development of transparent antibacterial and antiviral coatings. This work presented a novel short peptide (NH₂-DOPA-(Phe)₂-(His)₆-OH) that can self-assemble and bind different metal Cu ions. These coatings released monovalent cuprous ions and hydrogen peroxide, providing the surfaces with significant antibacterial effects and antiviral properties.

Antifungal systems were also approached. Pimentel et al. evaluated the antifungal activity against *C. albicans* of silver-containing microcrystals such as silver vanadate (α -AgVO₃), silver tungstate (α -Ag₂WO₄), and silver molybdate (β -Ag₂MoO₄). The biocompatibility was tested by an interestingly new three-dimensional coculture model involving three cell lines. Based on the biocompatibility and antifungal findings, α -Ag₂WO₄ was the most promising material for treating *C. albicans* infection.

Layer-by-layer (LbL) assembly is a very versatile and significantly inexpensive approach to forming thin films. In this field, Miranda et al. assembled by LbL an extracellular matrix protein fibronectin, as an anionic polymer, with poly-L-lysine (PLL), as a cationic polymer. The fibronectin component was able to create a biomimetic interface, which enhanced the cell attachment of fibroblasts. In addition, the presence of PLL conferred to the films remarkable antibacterial properties against Gram-positive strains.

Cell-based drug delivery has been highlighted as a system to facilitate the transfer of substances to specific organs and tissues with elevated efficiency. In that sense, Litvinova et al. demonstrated that hMSCs can be successfully impregnated with synthetic CaCO₃ microcapsules without damaging cell structural integrity and mobility. This study established the possibility of using hMSCs as a carrier for a targeted drug delivery system.

REFERENCES

- Madsen, H. J., Meguid, R. A., Bronsert, M. R., Dyas, A. R., Colborn, K. L., Lambert-Kerzner, A., et al. (2021). Associations Between Preoperative Risks of Postoperative Complications: Results of an Analysis of 4.8 Million ACS-NSQIP Patients. *Am. J. Surg.* [Epub ahead of print]. doi:10.1016/j.amjsurg.2021.11.024
- Major, M. R., Wong, V. W., Nelson, E. R., Longaker, M. T., and Gurtner, G. C. (2015). The Foreign Body Response: At the Interface of Surgery and Bioengineering. *Plast. Reconstr. Surg.* 135 (5), 1489–1498. doi:10.1097/PRS.0000000000001193
- Ratner, B. D., and Zhang, G. (2020). "A History of Biomaterials," in *Biomaterials Science* (Cambridge, MA, United States: Elsevier), 21–34. doi:10.1016/B978-0-12-816137-1.00002-7
- Singh, R., Bathaei, M. J., Istif, E., and Beker, L. (2020). A Review of Bioresorbable Implantable Medical Devices: Materials, Fabrication, and Implementation. *Adv. Healthc. Mater.* 9 (18), 2000790. doi:10.1002/adhm.202000790

Finally, Picker et al. contributed with a review on the use of prokaryotic collagen-like proteins to both study the biology of conventional collagen and develop a new biomaterial platform. Moreover, the authors described the application of Scl2 protein in thromboresistant coating for cardiovascular devices, scaffolds for bone regeneration, chronic wound dressing, and matrices for cartilage regeneration.

In conclusion, this Research Topic provides an overview of recent advances in biomacromolecules systems used for the release of different kinds of therapeutic agents. Moreover, the exploration of the use of innovative carriers transfers substances such as cells and proteins.

AUTHOR CONTRIBUTIONS

All authors listed have made a substantial, direct, and intellectual contribution to the work and approved it for publication.

FUNDING

This work was partly supported by FONDECYT– Chile (Grant Number 11180395).

ACKNOWLEDGMENTS

The editors appreciate the contributions of all authors to this Research Topic, the constructive comments of all the reviewers, and the editorial support from Frontiers throughout the publication process.

Conflict of Interest: The authors declare that the research was conducted in the absence of any commercial or financial relationships that could be construed as a potential conflict of interest.

Publisher's Note: All claims expressed in this article are solely those of the authors and do not necessarily represent those of their affiliated organizations, or those of the publisher, the editors and the reviewers. Any product that may be evaluated in this article, or claim that may be made by its manufacturer, is not guaranteed or endorsed by the publisher.

Copyright © 2022 Lopes, Tabary and Hernandez-Montelongo. This is an open-access article distributed under the terms of the Creative Commons Attribution License (CC BY). The use, distribution or reproduction in other forums is permitted, provided the original author(s) and the copyright owner(s) are credited and that the original publication in this journal is cited, in accordance with accepted academic practice. No use, distribution or reproduction is permitted which does not comply with these terms.