



OPEN ACCESS

EDITED AND REVIEWED BY
Michael Albert Thomas,
University of California, Los Angeles,
United States

*CORRESPONDENCE

Jingjing Liu
liujingjing198631@126.com
Jingliang Cheng
fccchengjl@zzu.edu.cn

[†]These authors have contributed
equally to this work

SPECIALTY SECTION

This article was submitted to
Cancer Imaging and
Image-directed Interventions,
a section of the journal
Frontiers in Oncology

RECEIVED 09 July 2022

ACCEPTED 12 August 2022

PUBLISHED 21 September 2022

CITATION

Liu J, Huang M, Ren Y, Xu M, Li Y,
Cheng J and Zhu J (2022)
Corrigendum: The evaluation of
zoomed echo-planar diffusion-
weighted magnetic resonance imaging
with two-dimensional spatial-selective
radiofrequency excitation
pulses in patients with
hilar cholangiocarcinoma.
Front. Oncol. 12:989870.
doi: 10.3389/fonc.2022.989870

COPYRIGHT

© 2022 Liu, Huang, Ren, Xu, Li, Cheng
and Zhu. This is an open-access article
distributed under the terms of the
Creative Commons Attribution License
(CC BY). The use, distribution or
reproduction in other forums is
permitted, provided the original
author(s) and the copyright owner(s)
are credited and that the original
publication in this journal is cited, in
accordance with accepted academic
practice. No use, distribution or
reproduction is permitted which does
not comply with these terms.

Corrigendum: The evaluation of zoomed echo-planar diffusion-weighted magnetic resonance imaging with two-dimensional spatial-selective radiofrequency excitation pulses in patients with hilar cholangiocarcinoma

Jingjing Liu^{1*†}, Mengyue Huang^{1†}, Yanan Ren¹, Man Xu¹,
Yinhua Li¹, Jingliang Cheng^{1*} and Jinxia Zhu²

¹Department of Magnetic Resonance Imaging, The First Affiliated Hospital of Zhengzhou University, Zhengzhou, China, ²MR Collaboration, Siemens Healthcare Ltd., Beijing, China

KEYWORDS

hilar cholangiocarcinoma, zoomed EPI, diffusion-weighted imaging, magnetic resonance imaging, image quality

A Corrigendum on

The evaluation of zoomed echo-planar diffusion-weighted magnetic resonance imaging with two-dimensional spatial-selective radiofrequency excitation pulses in patients with hilar cholangiocarcinoma

by Liu J, Huang M, Ren Y, Xu M, Li Y, Cheng J and Zhu J (2022). *Front. Oncol.* 12:816008. doi: 10.3389/fonc.2022.816008

In the published article, there was an error in [Table 1](#) as published. The acquisition time of z-EPI DWI in [Table 1](#) was displayed as “3min45s”. The correct statement is “3min18s~6min11s (depending on breathing pattern)”. The acquisition time of axial T₂WI in [Table 1](#) was displayed as “2min50s”. The correct statement is “3min44s~4min28s (depending on breathing pattern)”. The corrected [Table 1](#) and its caption “Magnetic resonance imaging parameters” appear below.

The authors apologize for this error and state that this does not change the scientific conclusions of the article in any way. The original article has been updated.

TABLE 1 Magnetic resonance imaging parameters.

Imaging parameters	c-EPI DWI	z-EPI DWI	Coronal T ₂ WI	Axial T ₂ WI	T ₁ WI/ DCE
TR [ms]	4500	2000	1400	3100	3.9
TE [ms]	56	61	67	87	1.89
FOV [mm ²]	350 × 292	230 × 120	360 × 360	380 × 380	380 × 309
Reconstructed voxel size (mm ³)	2.2 × 2.2 × 5	1.5 × 1.5 × 5	1.4 × 1.4 × 5	1.2 × 1.2 × 5	0.7 × 0.7 × 3
Acquisition matrix	158 × 121	154 × 50	256 × 256	320 × 320	288 × 187
Slice thickness [mm]	5	5	5	5	3
b-values [s/mm ²]	50/800	50/800	/	/	/
Parallel acceleration factor	2	2	3	2	2
Bandwidth (Hz/px)	1978	1248	698	710	1090
Respiration control	Free-breathing	Trigger	Breath-hold	Trigger	Breath-hold
Acquisition time	1min52s	3min18s~6min11s (depending on breathing pattern)	28s	3min44s~4min28s (depending on breathing pattern)	17s

Publisher's note

All claims expressed in this article are solely those of the authors and do not necessarily represent those of their affiliated organizations, or those of the publisher, the editors and the

reviewers. Any product that may be evaluated in this article, or claim that may be made by its manufacturer, is not guaranteed or endorsed by the publisher.