



OPEN ACCESS

EDITED AND REVIEWED BY
Francisco Ruben Carvalho Chaigneau,
Virginia Tech, United States

*CORRESPONDENCE
Chunhe Wan
✉ chunhewan@126.com

[†]These authors have contributed equally to this work

RECEIVED 05 December 2024
ACCEPTED 10 December 2024
PUBLISHED 13 January 2025

CITATION
Chen S, Chen Y, Zhang M, Zhang W, Fu H, Huang Y, Cheng L and Wan C (2025) Corrigendum: Specific detection of duck adeno-associated virus using a TaqMan-based real-time PCR assay. *Front. Vet. Sci.* 11:1539901. doi: 10.3389/fvets.2024.1539901

COPYRIGHT
© 2025 Chen, Chen, Zhang, Zhang, Fu, Huang, Cheng and Wan. This is an open-access article distributed under the terms of the [Creative Commons Attribution License \(CC BY\)](#). The use, distribution or reproduction in other forums is permitted, provided the original author(s) and the copyright owner(s) are credited and that the original publication in this journal is cited, in accordance with accepted academic practice. No use, distribution or reproduction is permitted which does not comply with these terms.

Corrigendum: Specific detection of duck adeno-associated virus using a TaqMan-based real-time PCR assay

Shuyu Chen^{1,2†}, YuYi Chen^{1,3†}, Mengyan Zhang^{1,3†},
Wenyu Zhang^{1,2†}, Huanru Fu^{1,2}, Yu Huang¹, Longfei Cheng¹ and
Chunhe Wan^{1*}

¹Fujian Key Laboratory for Avian Diseases Control and Prevention, Fujian Academy of Agricultural Sciences, Institute of Animal Husbandry and Veterinary Medicine, Fujian Animal Diseases Control Technology Development Centre, Fuzhou, China, ²School of Life Sciences, Fujian Agriculture and Forestry University, Fuzhou, China, ³College of Animal Sciences, Fujian Agriculture and Forestry University, Fuzhou, China

KEYWORDS

duck adeno-associated virus, DAAV, TaqMan, qPCR, epidemiological surveillance

A Corrigendum on

Specific detection of duck adeno-associated virus using a TaqMan-based real-time PCR assay

by Chen, S., Chen, Y., Zhang, M., Zhang, W., Fu, H., Huang, Y., Cheng, L., and Wan, C. (2024). *Front. Vet. Sci.* 11:1483990. doi: 10.3389/fvets.2024.1483990

In the published article, there was an error in [Table 2](#) as published. The Ct value of Inter-assay CVs should be changed as,

15.52, instead of 15.46;

22.31, instead of 22.24;

28.99, instead of 28.90.

The corrected [Table 2](#) and its caption appear below.

The authors apologize for this error and state that this does not change the scientific conclusions of the article in any way. The original article has been updated.

Publisher's note

All claims expressed in this article are solely those of the authors and do not necessarily represent those of their affiliated organizations, or those of the publisher, the editors and the reviewers. Any product that may be evaluated in this article, or claim that may be made by its manufacturer, is not guaranteed or endorsed by the publisher.

TABLE 2 Coefficient of variation of the TaqMan-qPCR assay.

| Serial number | Copies | Intra-assay CVs | | | Inter-assay CVs | | |
|---------------|--------------------|-----------------|--------------------|--------|-----------------|--------------------|--------|
| | | Ct | Standard deviation | CV (%) | Ct | Standard deviation | CV (%) |
| 1 | 2.91×10^6 | 15.46 | 0.12 | 0.76 | 15.52 | 0.10 | 0.62 |
| 2 | 2.91×10^4 | 22.24 | 0.09 | 0.41 | 22.31 | 0.15 | 0.67 |
| 3 | 2.91×10^2 | 28.90 | 0.21 | 0.72 | 28.99 | 0.42 | 1.42 |